

Annex of the Certificate (Page 1/2)


Scope of Accreditation

 Kalibrasyon TS EN ISO/IEC 17025 AB-0019-K	ESİT Elektronik Sistemler İmalat Ve Ticaret Ltd. Şti.	
	Accreditation Nr : AB-0019-K Revision : 01 Tarih: 06-Nisan-2007	
	Calibration Laboratory	
	Address Nişantepe Mah. No:8 Alemdağ Ümraniye 34794 İSTANBUL / TÜRKİYE	Phone : 0 216 585 18 18 Fax : 0 216 304 64 10 E-Mail : esit@esit.com.tr Website : www.esit.com.tr

Measured Quantity Instrument or Gauge	Range	Measurement Conditions	Best Measurement Capability Expressed as an Expanded Uncertainty (k=2) (±)	Açıklamalar
MASS	200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg 20 kg 50 kg		0,06 mg 0,08 mg 0,1 mg 0,12 mg 0,16 mg 0,2 mg 0,25 mg 0,3 mg 0,5 mg 1 mg 2,5 mg 5 mg 10 mg 25 mg 50 mg 100 mg 250 mg	Class F2 weights
Conventional mass	1mg 2mg 5 mg 10 mg 20 mg 50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg 20 kg 50 kg 100 kg 200 kg 500 kg 1000 kg		0,06 mg 0,06 mg 0,06 mg 0,08 mg 0,1 mg 0,12 mg 0,16 mg 0,2 mg 0,25 mg 0,3 mg 0,4 mg 0,5 mg 0,6 mg 0,8 mg 1 mg 1,6 mg 3 mg 8 mg 16 mg 30 mg 80 mg 160 mg 300 mg 800 mg 1,6 g 3 g 8 g 16 g	M1 Class weights

Annex of the Certificate (Page 2/2)

Scope of Accreditation

 Kalibrasyon TS EN ISO/IEC 17025 AB-0019-K	<p style="text-align: center;">ESİT Elektronik Sistemler İmalat Ve Ticaret Ltd. Şti.</p> <p style="text-align: center;">Akreditasyon No: AB-0019-K Revizyon No: 01 Tarih: 06-Nisan-2007</p>
--	---

Measured Quantity Instrument or Gauge	Range	Measurement Conditions	Best Measurement Capability Expressed as an Expanded Uncertainty (k=2) (\pm)	Açıklamalar
Non Automatic Weighing Machines	up to 50 kg	with F1 Class weights	$1 \cdot 10^{-5}$	Calibration performed by the Customers' Premises and according to Quality Management Handbook and Instructions
	up to 120000 kg	with M1 Class weights	$2 \cdot 10^{-4}$	

The personel whose names are given below are found satisfactory to conduct the calibration activities and to prepare the related calibration certificates.

Calibration of Non Automatic Weighing Machines:

Hilal KAYA
Yahya SARAÇ
Cemal ŞENER

Calibration of mass:

Hilal KAYA
Nurcan TAÇ

End of Scope

Atakan BAŞTÜRK
Secretary General