


Annex of the Certificate (Page 1/1)

Scope of accreditation

 TS EN ISO/IEC 17025 06016-01	HERAEUS ELECTRO-NİTE TERMOTEKNİK SAN. VE TİC. A.Ş. KALİBRASYON LABORATUVARI	
	Accreditation Number: 06016-01 Revision Number: 0 Date: 30 March 2005	
	Calibration Laboratory	
	Address : Orhan Işık Caddesi No: 9 Organize San. Bölgesi 06930 Sincan Ankara / TÜRKİYE	Phone : +90 0312 267 08 88 Fax : +90 0312 267 08 87 E-mail : eninfo@electro-nite.com.tr Website : www.electro-nite.com

Measured Quantity Instrument or Gauge	Range	Measurement Conditions	Best Measurement Capability Expressed as an Expanded Uncertainty(k=2) (±)	Remarks
ELECTRICAL Temperature Indicators				
S type Thermocouple	400°C–1760°C		0,6 °C	Calibration by electrical simulation using calibrator
K type Thermocouple	200°C–1370°C		0,4 °C	
R type Thermocouple	400°C–1760°C		0,6 °C	
B type Thermocouple	400°C–1820°C		1,1 °C	
J type Thermocouple	0,1°C–1200°C		0,3 °C	

End of Scope

The personel whose names are given below are found satisfactory to conduct the calibration activities and to prepare the related calibration certificates.

Calibration of temperature indicators:

-Tayfun ŞİMŞEK

-Murat ÖZ

22 April 2005

Atakan BAŞTÜRK
Secretary General